



www.IetLtd.com Proudly serving laboratories worldwide since 1979

CALL +847.913.0777 for Refurbished & Certified Lab Equipment

Alltech 590

Unit Part Ntlmber:	590508
Description:	Elite Micro Degasser 1ml-/min, 4 channel
Unit Serial Number:	590-0806-0669

TEST DATA

Pumpdown

Starting	750.0	mmHg
Ending	69.8	mmHg
Time	< 1	minutes

RPM

Starting	40.7
Ending	39.5
Min	39.5
Max	40.7

Vacuum (mmHg)

Starting	49.9
Ending	49.9
Min	49.9
Max	50.0

Validation Output: 251 mv

Specifications

Parameter	Elite Degassing System	Select Degassing System
Pump Down Time	10 Minutes (max)	5 minutes @ 400 RPM
Pump RPM	20 to 100 RPM (load dependent)	60 RPM (after pump down)
Vacuum after PD	50 mmHg +/-1 mmHg	<60 mmHg
Validation Output	5 mv per mmHg + / -1% of reading	N/A
-	+/-10 mv	-



www.IetLtd.com Proudly serving laboratories worldwide since 1979

CALL +847.913.0777 for Refurbished & Certified Lab Equipment